

## **Certification of Accuracy**

**Terriss Models:**

**Date:** June 25, 2004

T-03-015      Calibrated Platinum Wire Bulb 0.01%

The product identified above, manufactured in the United States, has been accurately calibrated using testing standards traceable to National Institute of Standards and Technology (formerly N. B.S.) to assure maximum accuracy and maintain tolerances within  $\pm 1/6^{\text{th}}$  of 1%. Accuracy is traceable to NIST Test No. 40674.

**TERRISS CONSOLIDATED INDUSTRIES**

Stephen W. Bodnovich  
Vice President